

# APPLICATION OF SUPERCRITICAL FLUID CHROMATOGRAPHY COUPLED TO TANDEM MASS SPECTROMETRY FOR PESTICIDE ANALYSIS IN FOOD

Cutillas, V; Murcia-Morales, M.; Martínez Galera, M.; Fernández-Alba, A. R.

European Union Reference Laboratory for Pesticide Residues in Fruit & Vegetables. University of Almeria, Agrifood Campus of International Excellence (ceiA3)  
04120, Almería, Spain. email: amadeo@ual.es

## INTRODUCTION

Supercritical fluid chromatography (SFC) uses carbon dioxide (CO<sub>2</sub>) as a component of the mobile phase. The CO<sub>2</sub> when is put through specific conditions of temperature (31,1°C) and pressure (73 bar) acts as a solvent. This type of mobile phase provides different separation than combination of water and an organic solvent.

A fast analytical method with supercritical fluid chromatography coupled to triple quadrupole mass spectrometry was validated to quantify 164 pesticides in three different matrices (tomato, orange and leek). A CO<sub>2</sub> gradient with MeOH as co-solvent was employed. Methanol was used also as post-column solvent (added after the chromatographic column). The duration of the method of analysis was 12 minutes eluting the last compound at the minute 7.3.

## EXPERIMENTAL



**System:** Nexera UC coupled to a Shimadzu LC-MS 8060

### SFC parameters:

- Injection volume: 2µL
- Flow rate: 1,5 mL/min
- Oven temperature: 40°C
- BPR pressure: 150 bar
- BPR Temperature: 50°C
- Column: Shim-Pack UC-X RP, 3µm 2.1x250mm<sup>2</sup>
- Mobile Phases:  
Modifier: MeOH 1mM HCOONH<sub>4</sub>  
Make up: MeOH 5mM HCOONH<sub>4</sub> 0.1% HCOOH
- Gradient time: 9 min + 3 min reequilibration

### MS parameters:


- Ion source: ESI
- Polarity: Positive and negative
- Schedule MRM software features
- Dwell time: 5 ms



Chromatogram of the 164 pesticides validated in the method spiked at the concentration of 10µg/Kg in tomato.


## RESULTS AND DISCUSSION

### Percentage of identified compounds




### Reproducibility


For identified compounds at 5µg/kg




### Matrix Effect



### Proquinazid




### Carbaryl



### Sensitivity of acidic/polar compounds

Comparison of the sensitivity for the same compounds in liquid and critical fluid chromatography.




### Real Samples

Fruits and vegetables



Number of samples analyzed	25
Samples with pesticides	18 (72%)
Range of pesticides detected per sample (LOQ: 5 µg/kg)	0 - 7
Total load range (µg/kg)	0 - 183
Most detected pesticides	Chlorantraniliprole (3) Cyprodinil (3) Fenpyroximate (3) Fluopyram (3)

### Number of pesticides detected




### Real Samples

Spices




Number of samples analyzed	50
Samples with pesticides	43 (86%)
Range of pesticides detected per sample (LOQ: 5 µg/kg)	0 - 23
Total load range (µg/kg)	0 - 5364
Most detected pesticides	Carbendazim (21) Imidacloprid (16) Acetamiprid (13)

### Number of pesticides detected




### Proficiency Test

Fipronil in eggs (JRC-GEEL 2017)



### Proficiency Test

Pesticides Residues in pear purée (FAPAS 2017)



## CONCLUSIONS

- Supercritical fluid chromatography facilitates the high flow rates providing short analysis times.
- Despite low injection volume (2µL) the developed SFC-MS/MS method allowed the identification of the majority of 164 target pesticides at the concentration of 5 µg/kg in tomato, orange and leek.
- The majority of the analytes showed no significant matrix effects. For 98% of the study pesticides in tomato, 85% in orange and 62% in leek, the suppression was lower than 20%.
- The absence of water provide better sensitivity of acidic/polar compounds. Furthermore, the use of 100% organic solvent improves ionization efficiency and increases sensitivity.